



# 强制检定证书

## MANDATORY VERIFICATION CERTIFICATE

证书编号 YLZ202403125  
Certificate No.

第 1 页, 共 3 页  
Page of

送检单位 连州市人民医院 (一门诊)  
Client

送检单位地址 连州市洋湄路5号  
Add. of Client

计量器具名称 口腔额面锥形束计算机体层摄影设备 (全景口腔部分)  
Description

型号/规格 RCT700  
Model/Type

制造厂 RAYSCAN  
Manufacturer

出厂编号 RA72235137  
Serial No.

设备管理编号  
Equipment No.

检定依据 JJG1101-2014 医用诊断全景牙科X射线辐射源检定规程  
Regulation

检定结论 合格  
Conclusion Qualified

检定日期 2024 年 02 月 29 日  
Date of Verification Y M D

依据检定规程, 被检仪器检定周期不超过 壹 年  
The verification period is within 1 year(s)

批准人 吴国光 吴国光  
Approved Signatory

核验员 陈上文 陈上文  
Reviewed by

检定员 付勇明 付勇明  
Verified by

检定专用章  
Stamp



扫一扫查真伪



## 说明

证书编号 YLZ202403125  
Certificate No.

### DIRECTIONS

第 2 页, 共 3 页  
Page of

1. 本中心是国家市场监督管理总局在华南地区设立的国家法定计量检定机构, 计量授权证书号是: (国) 法计(2022) 01043号、(国) 法计(2022) 01032号。本中心质量管理体系符合JJF1069-2012《法定计量检定机构考核规范》的要求。

This laboratory is the National Legal Metrological Verification Institution in southern China set up by the State Administration for Market Regulation under authorization certificates No.(2022)01043 & (2022)01032. The quality system is in accordance with JJF1069-2012 Rules for the Examination of the Service of Legal Metrological Verification.

2. 本中心所出具的数据均可溯源至国家计量基准和国际单位制(SI)。

All data issued by this laboratory are traceable to national primary standards and International System of Units (SI).

3. 检定地点、环境条件:

Place and environmental conditions of the verification:

地点 委托方口腔科	温度 18.8 °C	相对湿度 60 %
Place	Temperature	R.H.

4. 本次检定所使用的主要计量标准器具:

Major standards of measurement used in the verification:

设备名称/型号规格/测量范围	编号	证书号/有效期/溯源单位	计量特性
Name of Equipment /Model/Type/Range	Serial No.	Certificate No./Due Date /Traceability to	Metrological Characteristic
X射线机多功能质量检测仪 Dosimetric Instruments for X-Ray /B-Piranha657/(0.1~999.9)mGy	CB2-15040191	DLj12023-07073 /2024-05-29 /国家计量院	$U_{rel}=3.2\% (k=2)$
全景牙科机性能检测模体 Dental Panorama Detection Phantom /RTI D-P/低对比度: (1.0~2.5)mm; 空间分辨力: (16~30) Lp/cm 低对比度分 辨力: (1.0~2.5) mm	D-P-15-0108	DLj12023-07076 /2024-05-15 /国家计量院	低对比度MPE: $\pm 10\%$ ; 空间分 辨力MPE: $\pm 15\%$ Low contrast MPE: $\pm 10\%$ ; Spatial resolution MPE: $\pm 15\%$

注: 1. 本证书检定结果只与受检定仪器有关。The results relate only to the items verified.

Note: 2. 未经本机构书面批准, 不得部分复制此证书。This certificate shall not be reproduced except in full, without the written approval of our laboratory.

3. “委托方”、“委托方联络信息”由委托方提供, “制造厂”、“型号规格”、“出厂编号”以及“设备编号”为仪器上标注, 委托方对上面内容如有异议, 须在收到证书后二十个工作日内提出。

The information **Client** and **Contact Information** are provided by client, and the **Manufacturer**, **Model/Type**, **Serial No.** and **Equipment No.** are marked on the items. Client shall submit any objection within 20 working days after receiving the certificate for the information above.

4. 本次检定日期视为发布日期。The verification date is the date of issue of the certificate.



## 检定结果 RESULTS OF VERIFICATION

证书编号 YLZ202403125  
Certificate No.

原始记录号 YLZ202403125  
Record No.

第 3 页, 共 3 页  
Page of

1. 外观及功能性检查:	合格
Check on Appearance and Performance:	Pass
2. 空气比释动能率	
Air Kerma Rate of Radition Output	
测量条件:	80 kV, 14 mA
Measuring Condition	
空气比释动能率 $K$ =	44.939 mGy/min [技术要求: $\leq 60$ mGy/min]
Air Kerma Rate	Technical Requirements
3. 辐射输出的质	
Half-value Layer of Radiation Output	
测量条件:	70 kV, 10 mA
Measuring Condition	
半价层 $HVL$ =	3.6 mmAl [技术要求: $\geq 2.1$ mmAl]
Half-value Layer	Technical Requirements
4. 管电压示值误差:	-4.1 % [技术要求: $\pm 10$ %]
Voltage of the Tube	Technical Requirements
5. 空间分辨力:	20 Lp/cm [技术要求: $\geq 20$ Lp/cm]
Spatial Resolution	Technical Requirements
6. 低对比度分辨力:	1 mm [技术要求: 可分辨模体中0.5 mm
Low Contrast Resolution	厚的铝膜内孔径为1 mm的圆孔]
	Technical Requirements

说明:

Note:

- 测量结果的扩展不确定度: 空气比释动能率 $U_{rel}=6\%$ , 管电压 $U_{rel}=3.0\%$ , ( $k=2$ )。  
Expanded uncertainty of measurement results.
- 本证书中给出的扩展不确定度依据 JJF1059.1-2012 《测量不确定度评定与表示》评定, 由合成标准不确定度乘以包含概率约为95%时对应的包含因子  $k$  得到。  
The expanded uncertainty given in this certificate is evaluated according to JJF1059.1-2012 "Evaluation and Expression of Uncertainty in Measurement", which is obtained by multiplying the combined standard uncertainty by the coverage factor  $k$  corresponding to the coverage probability of about 95%.